

Medical Maintenance Codes

MEDICAL EQUIPMENT REPAIRMAN'S EVALUATION CODES						
Repairman's Code	Rating	Repairman's Code With Problem In:				
		<i>Electrical</i>	<i>Mechanical</i>	<i>Hydraulic</i>	<i>Plumbing</i>	<i>Surface</i>
A	Excellent	B	C	D	E	F
G	Good	H	I	J	K	L
M	Fair	N	O	P	Q	R
S	Item has been reconditioned					
T	Item is unserviceable and must be replaced					
U	Item is operational, but does not meet safety requirements (Electrical or Mechanical)					
V	Item does not perform to acceptable technical standards					
W	Item is not in use					
X	Parts are no longer available					
Y	Item is in repairable condition					
Z	Salvage - excessive maintenance or physical condition dictates against further use under any condition					

RELATIONSHIP BETWEEN EQUIPMENT STATUS CODE AND THREE-YEAR EQUIPMENT REPLACEMENT SCHEDULE				
Equipment reported as:		1 ST Year	2 ND Year	3 RD Year
Repairman's Code	Life Remaining	Calculated % of Life Remaining		
T,U,V,W,X, or Z	N/A	100		
M,N,O,P,Q,R,S, or Y	0	<40	>40	
	1	<25	26-50	51-75
	2	<20	21-30	31-40
	>2	<15	16-30	31-40
G,H,I,J,K, or L	0			<50
	1			<51
	2			<31
	>2			<20

Maintenance Code	DLA Code	Title
A-L	A	Serviceable (issuable without qualification)
M-R	B	Serviceable (issuable with qualification)
Y	E,F	Unserviceable (reparable/limited restoration)
U,V,X	H	Unserviceable (Condemned)

Medical Maintenance Codes

T,Z	S	Unserviceable (Scrap)
------------	----------	------------------------------